



T.C. SAĞLIK BAKANLIĞI
HALK SAĞLIĞI GENEL MÜDÜRLÜĞÜ
Çevre Sağlığı Dairesi Başkanlığı
Yetkili Laboratuvar



Elektrikli Isıtıcı Patojen Hava Filtresi Airborne Virucidal Efficacy Test Report against SARS-CoV-2 (COVID-19)

This report was prepared by ANTİMİKROP AR-GE ve
BİYOSİDAL ANALİZ MERKEZİ

Nasuh Akar Mah. Süleyman
Hacıabdullahoğlu Cad. No: 37/1
Çankaya/Ankara for

**MCC Medikal Özel Sağlık Hizmetleri Endüstriyel Makine
Bilgisayar Elektronik Arge Danışmanlık Tarımsal Analiz
Laboratuvarı Sanayi ve Ticaret Ltd. Şti.**

*This report cannot be partially copied and reproduced without the written permission of the laboratory, and cannot be used outside the relevant institutions.
Unsigned reports are invalid.
This report is only valid for the sample whose information is above (delivered to our laboratory).
Our laboratory does not perform the sampling process, and our laboratory is not responsible for errors caused by sampling or problems that may occur due to the sample not representing the whole.
This report cannot be used for advertising purposes.
(*This experiment is not within the scope of accreditation.*


Customer : MCC Medikal Özel Sağlık Hizmetleri Endüstriyel Makine Bilgisayar Elektronik Arge Danışmanlık Tarımsal Analiz Laboratuvarı Sanayi ve Ticaret Ltd. Şti.

Contact : muratcanpolat01@gmail.com, 0505 858 2526

Address : Akdeniz Üniversitesi Dumlupınar Bulv. Uluğbey Ar-ge Binası Bodrum Kat No 1/L Konyaaltı/Antalya

Report Date : 15.11.2021

Report No : R-21-0333



4. RESULTS

4.1. Elektrikli Isıtıcı Patogen Hava Filtresi Test Results

TEST	Test Material	Log TCID ₅₀	Result	
Virus Titration after Control Test period*	Virus Solution (Control)	7,75	-	
Virus Titration after Test period*	Virus Solution (Test)	7,5	-	
Virucidal Air Disinfection Test 80 minutes	Control	5	R= Log K-Log T R=2,37	99,57%
	Test	2,63		

